

STANDARDS & QUALIFICATIONS TRACKER

TEST TYPE	INDUSTRY	TEST STANDARD	PRINT TECHNOLOGY	MATERIAL	PASS / FAIL
Skin Irritation	Medical	ISO 10993-10 (2013)	MJF	BASF Ultrasint TPU01	Pass
Skin Sensitivity	Medical	ISO 10993-10 (2013)	MJF	BASF Ultrasint TPU01	Pass
Hemolysis	Medical	ISO 10993-4	MJF	PA11	Pass
Cytotoxicity	Medical / Dental	ISO 10993-5 (2009)	MJF	PA11	Pass
Skin Irritation	Medical	ISO 10993-10 (2013)	MJF	PA11	Pass
Flammability	Automotive	FMVSS 302	MJF	PA12	Pass
Antibacterial Activity	Medical / Dental	ISO 22196 (2011)	MJF	PA12	Pass
Aquatic / Acute Toxicity	Medical	ISO 11348-3	MJF	PA12	Pass
Cytotoxicity	Medical	ISO 10993-5 (2009)	MJF	PA12	Pass
Intradermal Reactivity (Irritation)	Medical	ISO 10993-10 (2013)	MJF	PA11	Pass
Skin Irritation	Medical	ISO 10993-10 (2013)	MJF	Lubrizol TPU	Pass
Cytotoxicity	Medical	ISO 10993-5 (2009)	MJF	Lubrizol TPU	Pass
Food Contact	Industrial	DS/EN1186-01:2002, DS/EN1186-03:2002, DS/EN1186-14:	SLS	PA12	Pass
Skin Irritation	Medical	ISO 10993-10 (2013)	SLS	PA12	Pass
Cytotoxicity	Medical	ISO 10993-5 (2009)	SLS	PA12	Pass
Cytotoxicity	Medical	ISO 10993-5 (2009)	SLS	RICOH PP S5500P	Pass
Skin Sensitivity	Medical	ISO 10993-10 (2013)	SLS	PA12	Pass

AMT is providing this information to assist customers. It is the responsibility of each customer to determine that its particular use of AMT's post-processing is safe and technically suitable to the customer's intended applications and consistent with the relevant regulatory requirements applicable to the customer's final product. The only warranties for AMT products and services are set forth in the express warranty statements accompanying such specific products and services. Nothing herein should be construed as constituting an additional warranty. AMT shall not be liable for technical or editorial errors or omissions contained herein.



Verified by **Dr. Konstantin Rybalcenko**
Global Head of R&D at AMT